

## UKCA Type Examination – Production Type

This is to certify that:

Perry Baromedical Corporation  
3750 Prospect Avenue  
Riviera Beach  
Florida  
33404  
USA

Holds Certificate Number:

UKCA 752578

In respect of:

**Monoplace Hyperbaric Chambers - SIGMA 34, SIGMA 36 and SIGMA 40.**

On the basis of our examination under the requirements of the Pressure Equipment (Safety) Regulations (SI 2016/1105) as amended by the Product Safety and Metrology etc (Amendment etc)(EU Exit) Regulations 2020 (SI 2020/676) for Conformity Assessment Procedure Module B - Type Examination - Production Type.

For and on behalf of BSI, an Approved  
Body for the above Regulation  
(Approved Body Number 0086):



Frank Lee, Product Certification Technical and Compliance Director

First Issued: 2022-08-04

Latest Issue: 2022-08-04

Effective Date: 2022-08-04

Expiry Date: 2027-08-03

Page: 1 of 2

...making excellence a habit.™

# UKCA Type Examination – Production Type

No. UKCA 752578

## Additional Information

### Limitations of Certificate Validity

#### Monoplace Hyperbaric Chambers

Model Nos. SIGMA 34, SIGMA 36 and SIGMA 40

Technical File Ref. No. TECH-903638 Rev. F

PS = 2.34 bar g

TS = 0°C to 38°C

Fluid Groups 1 and 2 (Oxygen Rich Compressed Air)

Main Materials of Construction – Acrylic Window, Aluminium and Stainless/Carbon Steels

First Issued: 2022-08-04

Latest Issue: 2022-08-04

Effective Date: 2022-08-04

Expiry Date: 2027-08-03

Page: 2 of 2

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.  
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.  
A member of BSI Group of Companies.