

# CERTIFICATE OF COMPLIANCE

**Certificate Number:** SGSNA/25/SUW/00131X

**Contract Number:** 801324  
**Certificate Project Number:** SUW-CERT250900182

**Certified Product:** Sigma and Sigma Elite  
**Trademarks:** SIGMA Series Hyperbaric Chambers  
**Model(s):** Power Module D-903617  
**Technical Data:** 120 – 240 V~, 50-60 Hz, 1.6A

**Certificate Holder:** Perry Baromedical Corporation  
3595 Fiscal Court, 33404, Riviera Beach, Florida, United States of America

*This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at [www.sgs.com](http://www.sgs.com) or using the QR code below. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with:*

ANSI/AAMI ES60601-1:2005 + C1:2009 + A2:2010 + A1:2012  
CAN/CSA C22.2 No. 60601-1:14

### Certification Condition

- 1.Documentation and markings used for evaluation was provided in English, other languages shall be verified by the receiving countries.
- 2.The equipment was tested with a standard US/CAN hospital grade detachable power cord. Appropriate detachable cords set shall be verified by the receiving countries.
- 3.The D-903617 power module is only to be used with Sigma system.
- 4.The end system must be adequately approved to it's applicable standards.
- 5.All risk management material is covered by the system RMF documentation.

Authorized by:

Effective date: 17 February 2026

Ana\_Ruas  
Certifier



### Certification Body

Connectivity & Products, a division of SGS North America Inc.  
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA  
t +1 770 570 1800 f +1 770 277 1240 [www.sgs.com](http://www.sgs.com)

